

Carbon Monoxide (CO) Module

Model 7591-06

Operation Instructions

- 1. Power On: Hold down power button on left side of module for 1 second.
- 2. Power Off: Hold down power button on left side of module until you see a light sequence of green, yellow, red. That indicates module is off.
- Charging: Plug USB-C into bottom of module and into power source. A blinking green light indicates module is charging.
- 4. Connecting to a Smart Station: Power on module by holding down the circle on the left for 1 second. A blue LED light will start to blink indicating Bluetooth® is on.
 - The module will appear in the list on the Smart Station. If not hit Refresh
 - Press desired module on screen to connect to Smart Station.

 A green check mark indicates a successful connection.
 - Press Done

Now you're ready to take measurements.

The OmniTrak™ Smart Station can be used in conjunction with any OmniTrak™ Module to provide immediate feedback as to the conditions in the immediate area. Take measurements, create reports, and analyze data provided to help improve conditions. This instrument is a handheld device, not to be worn on the body, or near an individual's head.

Dispose of battery in accordance with local rules and regulations.

What's Included in the Box

- Device
- USB-C Cable

NOTE: All modules, by default, are shipped in 'Sleep Mode' to protect the battery. It will not be possible to power on the Module, using the side button, until 'Sleep Mode' is disabled. To disable, plug the module into a power source and charge the module for a few minutes. This process is required only once.

Carbon Monoxide (CO) Module Model 7591-06

Carbon Monoxide Sensor Specifications

Specifications

Sensor Type Electrochemical Range 0- 400 ppm Accuracy 15% + 2 ppm

Resolution 0.1 ppm
*Response Time 45 seconds

Measurement specifications apply at ambient conditions of 21 +/- 5 °C temperature, 98.6 +/- 5 kPa pressure, and 50 +/- 10% relative humidity.

AC and DC Power Requirements	
AC/DC Power Supply	100 to 240 VAC, 50/60 Hz, 0.3A
Device Input Power	5 VDC, 2A, 10W

Environmental/Installation Requirements	
3000 meters (10,000 feet)	
2	
I	
0-30 °C dew point	
-20-35 °C dew point	
0% - 95% (non-condensing)	

Compliance

FCC ID: IC ID: 28101-OTMODULES

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:

- 1. This device may not cause harmful interference.
- This device must accept any interference received, including interference that may cause undesired operation.

Per FCC 15.21 Changes or modifications not expressly approved by TSI will void the product warranty and may void the user's authority to operate the equipment.

Compliance

This device contains license-exempt transmitter(s)/receiver(s) that comply with Innovation, Science and Economic Development Canada's license-exempt RSS(s). Operation is subject to the following two conditions:

- subject to the following two conditions:

 1. This device may not cause interference.
- This device must accept any interference, including interference that may cause undesired operation of the device.

To satisfy RF exposure requirements, this device and its antenna may be operated near a persons body and the extremities, but must be operated with a separation distance of at least 20 cm from all person's head.

Pour satisfaire aux exigences d'exposition RF, cet appareil et son antenne peuvent être utilisés à proximité du corps et des extrémités d'une personne, mais doivent être utilisés avec une distance de séparation d'au moins 20 cm de la tête de toute personne.

^{*(}Typical) time to 90% of final value

Carbon Monoxide (CO) Module Model 7591-06

\wedge

WARNING

- DO NOT attempt to replace the battery in this instrument.
- DO NOT abuse the battery in any way as the battery may rupture or catch fire.
- DO NOT use a substitute or nonrechargeable battery in this instrument.
- DO NOT short-circuit, incinerate, dismantle or mutilate Lithium ion batteries
- . DO NOT expose to water or heat.
- DO NOT use any battery which shows signs of damage, such as bulging, swelling, a swollen plastic wrap, liquid in the plastic wrap, etc.



Knowledge Beyond Measure.

TSI Incorporated - Visit our website www.tsi.com for more information.

USA Tel: +1 800 874 2811 India Tel: +91 80 67877200 UK Tel: +44 149 4 459200 China Tel: +86 10 8219 7688 France Tel: +33 1 41 19 21 99 Singapore Tel: +65 6595 6388

Germany Tel: +49 241 523030

P/N 5003097 Rev C ©2024 TSI Incorporated Printed in U.S.A.



Specifications are subject to change without notice.

Wi-Fi is a registered trademark of Wi-Fi Alliance®.

The Bluetooth® word mark and logos are registered trademarks owned by Bluetooth SIG, Inc. and any use of such marks by [licensee name] is under license. Other trademarks and trade names are those of their respective owners.

TSI, and the TSI logo are registered trademarks of TSI Incorporated in the United States and may be protected under other country's trademark registrations.

6176978660