

Vibrating Orifice Aerosol Generator

Model 3450:RO-3450

Standard maintenance and calibration service for the Vibrating Orifice Aerosol Generator Model 3450 consists of these procedures:

- 1) Review written instructions or explanation of problem from customer
- 2) Initial inspection
- 3) Update manual as necessary
- 4) Clean instrument
- 5) Inspect and replace as needed: filters, tubing, and O-rings
- 6) Verify frequency and voltage accuracy
- 7) Final electrical checkout and performance verification

Standard service includes minor repairs and replacement parts. Major repairs and replacement parts will be charged for parts and labor. TSI will contact the customer for approval before making any major repairs.

Turnaround time for standard service is approximately two weeks after receipt of the instrument. TSI service work is backed by a 90 day service warranty.

TSI Incorporated – 500 Cardigan Road, Shoreview, MN 55126 U.S.A

USA Tel: +1 800 874 2811

UK Tel: +44 149 4 459200

France Tel: +33 491 95 21 90

Germany Tel: +49 241 523030

India Tel: +91 80 41132470

China Tel: +86 10 8260 1595

E-mail: particle@tsi.com

E-mail: tsiuk@tsi.com

E-mail: tsifrance@tsi.com

E-mail: tsigmbh@tsi.com

E-mail: tsi-india@tsi.com

E-mail: tsibeijing@tsi.com

Website: www.tsi.com

Website: www.tsiinc.co.uk

Website: www.tsiinc.fr

Website: www.tsiinc.de



TRUST. SCIENCE. INNOVATION.

Contact your local TSI Distributor or visit our website www.tsi.com for more detailed specifications.